



Consumer and
Corporate Affairs Canada

Consommation
et Corporations Canada

Legal Metrology
Branch

Direction de la
Métrologie légale

Holland Avenue
Ottawa, Ontario
K1A 0C9

Avenue Holland
Ottawa, Ontario
K1A 0C9

Your file *Votre référence*

Our file *Notre référence*

Telephone: (613)952-0635

File N°: 06565
Project: AP-EL-93-0016

Schlumberger Industries
Sangamo Electricity Division
215 Laird Drive
Toronto, Ontario
M4G 2X1
Attention: Mr. Marc Purc

1993-08-12

Dear Mr. Purc:

This is in response to your recent notification given in accordance with the Electricity and Gas Inspection Regulations, Section 13, regarding devices approved under Notice of Approval AE-0338. It is noted that these devices are to be modified as follows:

These meters may be equipped with a 4 dial/digit energy register and a x10 multiplier, with the meter constant Kh displayable in the alternate mode along with the identifier code "11". Because the multiplier will apply to all displayed quantities, the constant Kh will be displayed as one-tenth (1/10) of the actual value shown on the meter nameplate.

These meters may also be equipped with an access door (as part of the demand reset mechanism) to permit battery replacement without breaking the meter seal.

I am advised by the Electrical Metrology Laboratory that their evaluation of the proposed modifications indicates that they do not adversely affect the performance of the meter. Accordingly, your notification is acknowledged and the modifications are accepted as being included within the above-noted approval. All recipients of the approval notice are being informed by copy of this letter.

Yours truly,

H.L. Fraser, Chief
Electricity and Gas

cc: Approvals Mailing List

Canada