



Department of consumer and corporate affairs / Ministère de la consommation et des corporations



**STANDARDS BRANCH - DIRECTION DES NORMES**

**NOTICE OF APPROVAL  
AVIS D'APPROBATION**

**E-18-1**

OTTAWA September 30, 1974

LEEDS & NORTHRUP TYPE "H" SPEEDOMAX MULTI-POINT  
STRIP CHART POTENTIOMETER RECORDER

This Notice is an addition to E-18 which otherwise remains in effect.

Notice E-18 covers the single pen version of the Speedomax "H" recorder. This Notice extends the approval to include the following multi-point versions:

<u>No. of Points</u>	<u>Designation</u>
1	300
2	302
3	303

The following additional optional features "g" are approved:

-304	12 seconds per point
-305	30 seconds per point
-306	60 seconds per point.

The standard amplifier will be transistor type 444508 which replaces type 101108.

The measuring circuit is now designated (-101) and is used for all d.c. mv ranges from 1 to 1000 mv.

All other features and designations remain the same.

Approval granted to:

Leeds & Northrup, Canada Ltd.,  
41 Constellation Court,  
Rexdale, Ontario.

J. L. Armstrong, P. Eng.,  
Chief, Standards Laboratory,

D. L. Smith, P. Eng.,  
Chief, Electricity & Gas Division,  
Metrology and Laboratory Services.

Ref: GL 1145-57/L2-240