DEPARTMENT OF TRADE AND COMMERCE

STANDARDS BRANCH

Files: SE-85-6-4 and SL-100-558

OTTAWA, June 16, 1966.

TO: DISTRICT INSPECTORS OF ELECTRICITY AND GAS

Sangamo Type "ED30" Polyphase Demand-Energy Meters

This is to advise that approval has been granted by the Standards Branch to the Sangamo type "ED30" 22-element wye horizontal polyphase demandenergy meter with 240-volt voltage coils for use on a 240/416 volt 4-wire wye service.

As this approval covers all the current ranges presently listed on Circular E-36 of May 10, 1966 under the $2\frac{1}{2}$ -Element Wye design, will you please add "240" to the 120 and 345 voltages already shown. This change is to be made on all copies of the circular in your possession.

Also on this circular are two typographical errors. The first is on the 'Current Ranges' line of the 2½-Element Wye where 5-40 should read ".5-40"; and the second is on the 'Current Ranges' line of the 2½-Element Delta where 5-37.5 should read ".5-37.5". Will you please make these corrections on all your copies.

K. Crver.

Chief, Electricity and Gas Division,

Standards Branch.

EAW/mee