



Consumer and  
Corporate Affairs Canada

Consommation  
et Corporations Canada

Legal Metrology  
Branch

Direction de la  
Métrologie légale

Holland Avenue  
Ottawa, Ontario  
K1A 0C9

Avenue Holland  
Ottawa, Ontario  
K1A 0C9

*Your file*    *Voire référence*

Legal Metrology Laboratories  
Telephone: (613) 990-8621

*Our file*    *Notre référence*

06565-C2-10  
April 24, 1986  
AML-E-18

Canadian General Electric  
1130 West Charest Blvd.  
QUEBEC, Québec  
G1N 2E2

Attn: Mr. Jean-Yves Bérubé, P. Eng.  
Engineering Department

Dear Mr. Bérubé:

This letter is in response to your letter dated April 21, 1986, regarding the modification to the type I-70 watt-hour meter registers, which are currently approved in Notices of Approval E-110 and E-184. I am advised that the Electrical Laboratory has evaluated the proposed modification (i.e.: increasing the thickness of the take-off gear from 0.020 inch to 0.060 inch to reduce gear tooth wear) and concluded that the thicker gear is satisfactory.

Accordingly, the modification is hereby accepted as being an immaterial one and hence included within the existing approvals.

All recipients of Notices of Approval under the Electricity and Gas Inspection Act are being informed of this decision by copy of this letter.

Yours truly

W.R. Virtue  
Chief

cc: Mailing List (Electrical Approvals)

Canada