



Consumer and
Corporate Affairs

Consommation et
corporations

Standards

Normes

**NOTICE OF APPROVAL
AVIS D'APPROBATION**

E-160

Ottawa, August 3, 1978

WESTINGHOUSE CANADA TYPES "D4S" AND D4A" METER REGISTER

This Notice of Approval covers register modifications as detailed below:

- a) The changes in the registers result from the standardization of the gear designs in order to facilitate production.

Both the 4 dial and 5 dial registers will have the same basic design; the 4 dial register will have the fifth dial, shaft, and gear removed.

- b) A slotted plastic disc which slips over the 5th dial shaft between the pointer and register face will replace the aluminum disc on 4/5 dial x 1 clock registers.
- c) The numbers are now applied to the drums of cyclometer register by the silk screening method.

Previously the numbers were molded into the drum.

In construction, size, dial diameters and appearance the modified registers are similar to the registers previously approved and are completely interchangeable with corresponding registers used on the "D-" line watt-hour meters since 1956.

Approval Granted to:

Westinghouse Canada Limited
Hamilton, Ontario

D. L. Smith,
Chief, Electricity and Gas Division
Legal Metrology Branch

Ref: G 6565-C3-10