



Consumer and
Corporate Affairs

Standards

Consommation et
corporations

Normes

**NOTICE OF APPROVAL
AVIS D'APPROBATION**

E-160-2

Ottawa, June 11, 1982

Approval Granted To:

Approbation accordée à:

Westinghouse Canada Inc.
Box 510
Hamilton, Ontario
L8N 3K2

W.R. Virtue

Acting Chief
Legal Metrology Laboratories
Legal Metrology Branch
Standards Building
Tunney's Pasture, Holland Ave.,
OTTAWA, Ontario
K1A 0C9

Chef intérimaire
Laboratoires de la métrologie légale
Direction de la métrologie légale
Edifice des normes
Parc Tunney, Avenue Holland
OTTAWA, Ontario
K1A 0C9

- 1 -

WESTINGHOUSE CANADA TYPES "D4S" AND "D4A"
METER REGISTERS.

This notice constitutes an amendment to
Notice of Approval E160-1 dated October
17, 1979.

Henceforth all new Westinghouse watt hour
meters will not be fitted with the
register described in Notice of Approval
E160-1

Meters fitted with this register and
sealed prior to June 1, 1982 may remain
in service, but when removed from service
for re-sealing, these registers must be
replaced.

Reference: G6565-C3-10

ENREGISTREURS DES COMPTEURS WESTINGHOUSE
CANADA DE TYPES "D4S" ET "D4A"

Le présent avis constitue une modification
de l'avis d'approbation E160-1 en date du
17 octobre 1979.

A compter d'aujourd'hui, aucun nouveau
wattheuremètre ne peut être muni de
l'enregistreur décrit dans l'avis
d'approbation E160-1.

Tout compteur déjà équipé de l'enregistreur
susmentionné et plombé avant le 1^{er} juin
1982 peut continuer à être utilisé.
Toutefois, dès que ces compteurs seront
retirés aux fins d'un nouveau plombage,
leur enregistreurs devront être remplacés.

Référence: G6565-C3-10