



Consumer and  
Corporate Affairs Canada

Consommation  
et Corporations Canada

Legal Metrology  
Branch

Direction de la  
Métrologie légale

Holland Avenue  
Ottawa, Ontario  
K1A 0C9

Avenue Holland  
Ottawa, Ontario  
K1A 0C9

Your file    Votre référence

Legal Metrology Laboratories  
Telephone: (613) 952-0640

Our file    Notre référence

O6565-S2-5  
October 28, 1986  
AML-E-23

Sangamo Canada  
(Division of Schlumberger Canada Ltd.)  
215 Laird Drive  
Toronto, Ontario  
M4G 3X1  
Attention: Mr. C. Gardiner

Dear Mr. Gardiner:

This is in response to your letter dated September 18, 1986, regarding the type K300 Demand Energy Meter currently approved in Notices of Approval E-181 and E-181-1, and your intention to make available as an option, the "Slave" input on the lower connector instead of the top connector and to identify such modified apparatus by marking the words "SLAVE INPUT ON LOWER CONNECTOR PIN G (+50V) PIN H (COMMON)" on the switch plate and by rotating the locating key on the external pulse cable plug 90° to prevent improper connection.

I am advised that, upon evaluation of the proposed modifications, the Electrical Laboratory has concluded that they may be regarded as insignificant in relation to the approval provisions of the Electricity and Gas Inspection Act. Accordingly, pursuant to the authority delegated to me under the said Act:

1. your letter is hereby acknowledged as satisfying the notification requirements of section 13, Part III of the Electricity and Gas Inspection Regulations, SOR/86-131, 16 January, 1986, and
2. the proposed alteration is hereby accepted, in accordance with the aforementioned section, as being an immaterial one and hence included within the approval.

All recipients of Notices of Approval under the Electricity and Gas Inspection Act are being informed of this decision by copy of this letter.

Yours truly,

W.R. Virtue  
Chief

cc: Mailing List (Electrical Approvals)

**Canada**