



Consumer and
Corporate Affairs Canada

Consommation
et Corporations Canada

Legal Metrology

Métrieologie légale

APPROVAL No. · N° D'APPROBATION

E-194-1

MAY 28 1985

NOTICE OF APPROVAL

Westinghouse Type D5S Single Phase Watt-hour Meter

Applicant: Westinghouse Canada Inc.
Box 510
Hamilton, Ontario
L8N 3K2

Manufacturer: Westinghouse Canada Inc.

This Notice of Approval is an extension of Notice of Approval E-194 to include meters rated 1.0 to 100 amperes, 120 and 240 volts.

Type of Device: Type D5S single phase watt-hour meter, self-contained, 2 wire.

Ratings: 1.0-100 amperes, 120 volts, 2 wire, 60 Hz, Kh 1.8, register ratio 55 5/9.

1.0-100 amperes, 240 volts, 2 wire, 60 Hz, Kh 3.6, register ratio 27 7/9.

Description: These meters are of the same basic design and performance characteristics as the meters approved under Notice of Approval E-194.

Approval: The apparatus specified herein has been duly approved under the provisions of the Electricity Inspection Act R.S.C 1970, chapter E-4, and the Electricity Meter Regulations C.R.C. 1978, chapter 561 for use in Canada under the general conditions of the said Regulations.

File Number: G6565-C3-7

Chief
Legal Metrology Laboratories

W.R. Virtue

AVIS D'APPROBATION

Watt-heuremètre monophasé Westinghouse de Type D5S

Requérant: Westinghouse Canada Inc.
Box 510
Hamilton, Ontario
L8N 3K2

Fabricant: Westinghouse Canada Inc.

Le présent avis d'approbation est une addition à l'avis d'approbation E-194 pour inclure les débits des compteurs 1.0 à 100 ampères, 120 et 240 volts.

Appareil: Wattheuremètre monophasé D5S, autonome, à 2 fils.

Caractéristiques nominales: 1.0-100 ampères, 120 volts, 2 fils, 60 Hz, constante Kh de 1.8, rapport d'enregistrement de 55 5/9.

1.0-100 ampères, 240 volts, 2 fils, 60 Hz, constante Kh de 3.6, rapport d'enregistrement de 27 7/9.

Description: Ces compteurs présentent des caractéristiques de conception et de rendement essentiellement identiques à celles des compteurs approuvés en vertu de l'avis d'approbation E-194.

Approbation: L'approbation est accordée en vertu des dispositions de la Loi sur l'inspection de l'électricité S.R.C. 1970, chapitre E-4, et du Règlement sur les compteurs électriques C.R.C.c. 1978, chapitre 561. L'emploi est autorisé au Canada sous réserve des conditions générales dudit règlement.

Numéro de dossier: G6565-C3-7

Chef
Laboratoires de la Métrieologie légale

Canada