



Consumer and  
Corporate Affairs Canada

Consommation  
et Corporations Canada

Legal Metrology  
Branch

Direction de la  
Métrologie légale

Holland Avenue  
Ottawa, Ontario  
K1A 0C9

Avenue Holiand  
Ottawa, Ontario  
K1A 0C9

Your file    *Voire référence*

Our file    *Notre référence*

Telephone: (613)952-0640

File N°: 06565  
Project: AP-EL-87-0038  
          AP-EL-87-0064  
          AP-EL-88-0014

Appalachian Technologies Corporation  
9109 Forsyth Drive  
Charlotte, North Carolina  
USA 28217  
Attn.: Mr. R.H. Austin

1988-12-15

Dear Mr. Austin,

This is in response to your recent notification given in accordance with the Electricity and Gas Inspection Regulations, Section 13, regarding devices approved under Notice of Approval E-238, E-238 Add 1 and E-245. It is noted that these devices are to be modified as follows:

- The 1/8 multiplexed display approved on the model EDR-1000 Electronic Polyphase Demand Meter Register (Notices of Approval E-238 and E-238 Add 1) and the 1/16 multiplexed display approved on the model SST-2430C Electronic Totalizing Pulse Recorder (Notice of Approval E-245) are interchangeable and may be used on either device.

I am advised by the Electrical Metrology Laboratory that their evaluation of the proposed modifications indicates that they may be regarded as immaterial in relation to the approval provisions of the Act. Accordingly, your notification is acknowledged and the modifications are accepted as being included within the above-noted approval. All recipients of the approval notice are being informed by copy of this letter.

Yours truly,

W.R. Virtue, Chief  
Legal Metrology Laboratories

cc: Approvals Mailing List

Canada