

DEPARTMENT OF TRADE AND COMMERCE
Electrical and Gas Standards Laboratories

Ottawa, September 25th, 1930.


District Inspectors of Electricity & Gas:

Re: Sangamo, type S.3, single phase,
two and three wire wathour meter.

Dear Sirs,- I beg to advise you that the single phase, 2 and 3 wire, type S.3, wathour meter, for use on 25 and 60 cycle current, manufactured by the Sangamo Electric Company of Canada, Limited, Toronto, Ontario, has been approved by the National Research Council and may be admitted to verification in Canada.

The type S.3 meter is built in the following capacities; 5, 10, 15, 25 and 50 amperes, and in voltages of standard rating, 115, 230, 460 and 575 and in frequencies of 25, 30 and 60 cycles. It is built in both 2 and 3 wire construction with either glass or metal covers, clock or cyclometer type registers.

Yours truly,



Director.

See NRC #134