

DEPARTMENT OF TRADE AND COMMERCE


Electrical and Gas Standards Laboratories

Ottawa, February 18th, 1932.

District Inspectors of Electricity & Gas:-

Sirs, - I beg to advise that the change in the location of the power factor switch from the side to the front of the 4L.2 and the WD.2 Lincoln Sangamo Demand meters has been approved by the National Research Council.

Yours truly,

  
Director.

JLS/RS.