

DEPARTMENT OF TRADE AND COMMERCE
Electrical and Gas Standards Laboratories

Ottawa, November 19th, 1932


District Inspectors of Electricity & Gas;-

Dear Sirs,- I beg to advise you that the type HC 2 element self-contained polyphase watt-hour meter for use on 3-phase 4-wire circuits without the use of transformers manufactured by the Sangamo Company Limited, Toronto, Ontario, has been approved by the National Research Council and may be admitted to verification in Canada.

This type of instrument will be built in capacities of 5, 10, 15, 25 and 50 amperes in the standard voltages and standard frequencies.

I am enclosing a copy of a testing specification provided by the Sangamo Company Limited, which will be, I think, helpful.

Yours truly,



Director.