

DEPARTMENT OF TRADE AND COMMERCE
ELECTRICITY AND GAS INSPECTION SERVICES

OTTAWA, Sept. 16, 1937.

TO DISTRICT INSPECTORS OF
ELECTRICITY AND GAS

Re: New Magnets for Types OO, CS, C-2,
C-5, CS-2 and CS-5 Westinghouse
Watt-hour Meters.

Dear Sirs:

The Canadian Westinghouse Company, Limited, has advised the National Research Council that they are supplying a new light-coloured finish on their permanent magnets for use in the above types of meters, and the National Research Council has advised the Company that they consider this change only an immaterial alteration and requires no further approval.

You are advised, therefore, that if you notice a change in this respect in the magnets of the above types of meters which are presented to you for verification, it will be in order for you to test and seal the meters without further question.

Yours truly,


Director.

JLS/mee