

File No. 241

M.R.C. 163  
Electricity

DEPARTMENT OF TRADE AND COMMERCE  
ELECTRICITY AND GAS INSPECTION SERVICES

OTTAWA, February 7, 1946.


TO DISTRICT INSPECTORS OF  
ELECTRICITY AND GAS

Re: Model "AW" Graphic Wattmeter -  
Esterline-Angus Company, Incorporated.

This is to advise you that the Model "AW" graphic wattmeter, manufactured by the Esterline-Angus Company, Inc., Indianapolis, Indiana, and sold in Canada by W. O. Taylor & Company, Ltd., Montreal, P. Q., has been approved by the National Research Council for use in Canada in the following capacities:

Amperes: 5/10

Volts: 100 to 750.



J. L. Stiver,  
Director.

JLS/ncc