



CANADA

DEPARTMENT OF TRADE AND COMMERCE
STANDARDS DIVISION

SD-EA.21

OTTAWA, May 23, 1950
(Superseding Circular Letter SD-EC.67,
dated January 12, 1950)

NOTIFICATION OF TYPE APPROVAL

The apparatus specified and illustrated herein has been duly approved by the Standards Division under the provisions of The Electricity Inspection Act, Chapter 22, 1928, and may be admitted to verification in Canada.

Apparatus Approved: Type "B" Cyclometer Register, for use on types "A1A", "A2A", "A1S", "A2S", and "FA" watt-hour meters, manufactured by the Packard Electric Company, Limited, St. Catharines, Ontario.

Rating of Apparatus: All approved ratings of the above-named types of meters.

Description: The construction of the type "B" cyclometer register is essentially the same as that of the type "B" clock register except for differences necessary in the construction of cyclometer dials. The plates, bosses, and gears may be made of anodized aluminum alloy or nickelled brass. This register will be supplied whenever a customer requests a complete new register for types "A", "A1" and "A2" watt-hour meters. It supersedes the present type "A" cyclometer register.

E. F. Power

E. F. Power,
Assistant Director (E. & G.),
Standards Division.

R. W. MacLean
R. W. MacLean,
Director,
Standards Division.

PACKARD TYPE "B" CYCLOMETER REGISTER

