



TRADE AND COMMERCE
CANADA

STANDARDS DIVISION

OTTAWA,..... March 19, 1953.

TYPE APPROVALBRISTOL CHART

The apparatus specified herein has been duly approved by the Standards Division under the provisions of The Electricity Inspection Act, Chapter 22, 1928, as amended, and may be admitted to verification in Canada.

Apparatus Approved: Roll Chart #R-0336, manufactured by The Bristol Company, Waterbury, Conn., U.S.A., and distributed in Canada by The Bristol Company of Canada Limited, 71-79 Duchess Street, Toronto 2, Ontario.

Description: This is a roll chart $12\frac{1}{4}$ inches wide with a calibrated scale of 11 inches. There are 120 feet of chart in a roll.

<u>Designation</u>	<u>Range</u>	<u>Time Base</u>	<u>Scale</u>
R-0336	0-10	2 inches/hour	Linear

E. F. Power

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean
R. W. MacLean,
Director,
Standards Division.

Ref: A-258A