



TRADE AND COMMERCE
CANADA

SD-EA.109

STANDARDS DIVISION

OTTAWA, June 9, 1953.

MODIFICATION OF TYPE APPROVAL

CANADIAN WESTINGHOUSE POLYPHASE WATTHOUR METERS
TYPES "CS-2" AND "CS-5"

The modification to the apparatus specified herein has been duly approved by the Standards Division under the provisions of The Electricity Inspection Act, Chapter 22, 1928, as amended, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: Types "CS-2" and "CS-5" Polyphase Watthour Meters, manufactured by the Canadian Westinghouse Company Limited, Hamilton, Ontario.

Rating of Apparatus:

Rated Amperes	2.5, 5, 10, 15, 25, 50
Maximum Amperes	7.5, 15, 30, 45, 75, 100
Voltage	115, 230, 460, 575
Frequency	60 cycles
Elements	2

Modification: These meters were originally approved under Circular NRC-98 of June 25, 1937. The modification consists of assigning maximum ampere ratings to these meters to conform with the maximum ratings approved for types "CA-2" and "CA-5" (see Circular NRC-164 of April 17, 1946).

E. F. Power

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean
R. W. MacLean,
Director,
Standards Division.