

TRADE AND COMMERCE
CANADA

SD-EA.142

STANDARDS DIVISION

OTTAWA, February 23, 1954.

MODIFICATION OF TYPE APPROVALCANADIAN WESTINGHOUSE TYPES "CA" AND "CS-1" WATTHOUR METERS

The modification to the apparatus specified herein has been duly approved by the Standards Division under the provisions of the Electricity Inspection Act, Chapter 94, R.S. 1952, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: Types "CA" and "CS-1" Single-Phase Watthour Meters, manufactured by the Canadian Westinghouse Company Limited, Hamilton, Ontario.

Rating of Apparatus: All previously approved ratings and capacities.

Modification: The modification consists in making the shaft of the light-load adjuster screw longer and fixing on the end of it a die-cast head. The head is locked in place on the shaft by means of a set screw. This arrangement is intended to make the setting of light load easier.

E. F. Power

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean

R. W. MacLean,
Director,
Standards Division.

Ref: A-189D