

TRADE AND COMMERCE  
CANADA

SD-EA.161

## STANDARDS DIVISION

OTTAWA.....October 29, 1954.

MODIFICATIONS OF TYPE APPROVALSANGAMO POLYPHASE WATTHOUR METER TYPE "S-3S"

The modifications to the apparatus specified herein have been duly approved by the Standards Division under the provisions of the Electricity Inspection Act, Chapter 94, R.S. 1952, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: Type "S-3S" Horizontal Polyphase Watthour Meter, manufactured by the Sangamo Company Limited, Leaside, Toronto 17, Canada.

Rating of Apparatus: All previously approved ratings.

Modifications: The main modification consists of alnico magnets in place of chrome steel magnets: see SD-EA.158 of September 29, 1954.

To accommodate the new magnet assembly, a die-cast grid will be used, and the register brackets will be slightly modified.

E. F. Power,  
Assistant Director (E&G),  
Standards Division.

R. W. MacLean,  
Director,  
Standards Division.

Ref: A-395