

TRADE AND COMMERCE
CANADA

SD-EA.171

STANDARDS DIVISION

OTTAWA, December 29, 1954.

TYPE APPROVALPACKARD TYPES "B1A" AND "B1S" WATTHOUR METERS

The apparatus specified herein has been duly approved by the Standards Division under the provisions of the Electricity Inspection Act, Chapter 94, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Types "B1A" and "B1S" Single-Phase Watthour Meters, manufactured by the Packard Electric Company Limited, St. Catharines, Ontario.

Rating of Apparatus: All previously approved ratings.

Description: These meters are exactly the same as the types "B1A" and "B1S" approved by earlier circulars except that the voltage ratings on the nameplates are multiples of 120 instead of 115.

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean,
Director,
Standards Division.

Ref: A-413