

TRADE AND COMMERCE
CANADA

SD-EA.193

STANDARDS DIVISION

OTTAWA,.....April 28, 1955.

TYPE APPROVALCANADIAN GENERAL ELECTRIC
TYPES "I-55A" AND "I-55S" WATTHOUR METERS

The apparatus specified herein has been duly approved by the Standards Division under the provisions of the Electricity Inspection Act, Chapter 94, R.S. 1952, and may be admitted to verification in Canada.


Apparatus Approved: Types "I-55A" and "I-55S" Single-Phase Watthour Meters, manufactured by the Canadian General Electric Company Limited, Quebec 8, P. Q.

Rating of Apparatus: As previously approved except for the voltage rating. (see Approval Circular SD-EA.189)

Description: These meters are identical to the same types approved under Circular SD-EA.189 except that the manufacturer is permitted to mark the nameplates with either a 240- or a 230-volt rating.

E. F. Power

E. F. Power,
Assistant Director (E&G),
Standards Division.



R. W. MacLean,
Director,
Standards Division.

Ref: A-436

