



TRADE AND COMMERCE
CANADA

STANDARDS DIVISION

OTTAWA, March 23, 1956.

TYPE APPROVAL

MINNEAPOLIS-HONEYWELL BROWN
"ELECTRONIK" RECORDERS WITH RETRANSMITTERS

The apparatus specified herein has been duly approved by the Standards Division under the provisions of the Electricity Inspection Act, Chapter 94, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Retransmitting slidewires for use with Models 153x10V, 153x11V and 153x12V Strip Chart "Elektronik" Recorders, manufactured by the Minneapolis-Honeywell Regulator Company Limited., Toronto 17, Ont.

Rating of Apparatus: See ratings for Models 153x10V, 153x11V, 153x12V on Approval Circular SD-EA.199, June 16, 1955.

Description:

This circular grants approval for the use of retransmitting slidewires with Models 153x10V, 153x11V, and 153x12V Strip Chart Recorders.

The slidewires used in the "Elektronik" Recorders are of two types: the "straight" slidewire and the "helical" slidewire. When one type is used for measuring, the other type is used for retransmitting. Thus, in the cases of Models 153x10V and 153x11V, since the measuring slidewire is of the "helical" type, the retransmitting one is a 250 ohm "straight" type. The Model 153x12V uses the "straight" type for measuring and a 250 ohm "helical" type is added when retransmitting is desired.

The inclusion of a retransmitting slidewire is indicated in the model designation by the letters B5 so that a typical model number would be 153X10V -X-20B5.

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean,
Director,
Standards Division.

