

TRADE AND COMMERCE  
CANADA

SD-EA.287

## STANDARDS DIVISION

OTTAWA, March 20, 1957.

MODIFICATION OF TYPE APPROVALFERRANTI TYPES "FMXA" AND "FMXS"  
WATTHOUR METERS

The modification to the apparatus specified herein has been duly approved by the Standards Division under the provisions of the Electricity Inspection Act, Chapter 94, R.S. 1952, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: Types "FMXA" and "FMXS" Single-Phase Watthour Meters, manufactured by Ferranti Electric Limited, Mount Dennis, Toronto 15, Ontario.

Rating of Apparatus: All previously-approved ratings.

Modification: This modification consists in fitting the meter disc with two anti-creep holes diametrically opposite on the disc instead of slots on the circumference as heretofore.

E. F. Power,  
Assistant Director (E&G),  
Standards Division.

R. W. MacLean,  
Director,  
Standards Division.

Ref: A-633

