



STANDARDS DIVISION

OTTAWA, March 7, 1958.

TYPE APPROVALSANGAMO JUMPER-TYPE CYCLOMETER REGISTER

The apparatus specified herein has been duly approved by the Standards Division under the provisions of the Electricity Inspection Act, Chapter 94, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Jumper-type Cyclometer Register, for use on the Type "EDA" single-phase demand-energy meters, manufactured by the Sangamo Company Limited, Leaside, Toronto 17, Ontario.

Rating of Apparatus: All approved ratings of the Type "EDA" demand-energy meters.

Description: The cyclometer register approved herein is the same register previously approved for use on the Sangamo Types "H", "S-2" and "CFA" watt-hour meters, and its use is extended to cover the Type "EDA" meter as well. The register pads or bosses on the meter grid of the Type "EDA" are reduced to compensate for the extra thickness of the cyclometer register. Spacers are required if a clock-type is to be used on meters built for the cyclometer register.

E. F. Power,
Assistant Director (L&G),
Standards Division.

R. W. MacLean,
Director,
Standards Division.

Ref: A-742

