

TRADE AND COMMERCE
CANADA

STANDARDS DIVISION

OTTAWA November 13, 1958.

MODIFICATION OF TYPE APPROVALCANADIAN GENERAL ELECTRIC
TYPES "I-50A" AND "I-55A" WATTHOUR METERS

The modifications to the apparatus specified and illustrated herein have been duly approved by the Standards Division under the provisions of the Electricity Inspection Act, Chapter 94, R.S. 1952, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: Types "I-50A" and "I-55A" Single-Phase Watthour Meters, manufactured by the Canadian General Electric Company Limited, Quebec, P. Q.

Rating of Apparatus: All approved capacities and ratings.

Modifications: The modifications are as follows:-

- 1) The provision of removable seal pans;
- 2) New means for locking the terminal block by means of a pin attached to the sealing clip, which eliminates the need for the screw through the seal pan;
- 3) Glass covers with removable cover rings which are held in place by spring sealing clips.

E. F. Power,
Assistant Director (SG),
Standards Division.

R. W. MacLean,
Director,
Standards Division.

Ref: A-790

CANADIAN GENERAL ELECTRIC
TYPES "I-50A" AND "I-55A" METER BASE



