



STANDARDS DIVISION

OTTAWA, March 16, 1959.

MODIFICATION OF TYPE APPROVALCANADIAN WESTINGHOUSE TYPE "D" CYCLOMETER REGISTER

The modification to the apparatus specified and illustrated herein has been duly approved by the Standards Division under the provisions of the Electricity Inspection Act, Chapter 94, R.S. 1952, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: Type "D" Cyclometer Register for use with types "DA" and "DS" watthour meters, manufactured by the Canadian Westinghouse Company Limited, Hamilton, Ontario.

Rating of Apparatus: These registers may be used with any approved rating of types "DA" and "DS" watthour meters.

Modification: These registers, which supersede the formerly-used registers, feature a chromate-treated aluminum take-off gear in place of gold-plated brass previously used. The new gear is dimensionally identical with the older type.

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean,
Director,
Standards Division.

Ref: A-659A

CANADIAN WESTINGHOUSE TYPE "D" CYCLOMETER REGISTER



