

TRADE AND COMMERCE
CANADA

S-EA.442

STANDARDS BRANCH

OTTAWA, March 22, 1960.

TYPE APPROVALCANADIAN WESTINGHOUSE TYPE "D" CLOCK REGISTER
WITH MULTIPLIER OF 10

The apparatus specified herein has been duly approved by the Standards Branch under the provisions of the Electricity Inspection Act, Chapter 94, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Type "D" Clock Register with multiplier of 10 for use with certain types "DA" and "DS" watt-hour meters, manufactured by the Canadian Westinghouse Company Limited, Hamilton, Ontario.

Rating of Apparatus: These registers may be used with suitable approved ratings of types "DA" and "DS" meters.

Description: The register is similar to that approved for use with the types "DAH" and "DSH" demand-energy meters and will differ only in that "Multiply by 10" will be printed on register dial face. The register ratio will be 277-7/9. Except for the register ratio, it is similar in construction to the register approved under SD-EA.405, May 22, 1959.

E. F. Power,
Assistant Director (E&G),
Standards Branch.

R. W. Maclean,
Director,
Standards Branch.

Ref: A-814B

