



TRADE AND COMMERCE
CANADA

STANDARDS BRANCH

OTTAWA, March 31, 1960.

TYPE APPROVAL

LEEDS & NORTHRUP "SPEEDOMAX" TYPE H MODEL S RECORDERS

The apparatus specified herein has been duly approved by the Standards Branch under the provisions of the Electricity Inspection Act, Chapter 94, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Type H Model S Strip-Chart "Speedomax" Potentiometer Recorders, manufactured by Leeds & Northrup, Canada, Ltd., 61 Industry Street, Toronto 15, Ontario.

Rating of Apparatus: See Approval Circular SD-EA.354, May 14, 1958 for details of rating and description of instrument nomenclature.

Description: These instruments are similar to the type H previously approved except that the standard cell and dry cell have been replaced by a constant voltage D.C. power supply of the Zener Diode variety. The external appearance and application are unchanged.

E. F. Power

E. F. Power,
Assistant Director (E&G),
Standards Branch.

R. W. MacLean
R. W. MacLean,
Director,
Standards Branch.

Ref: A-240B and A-660

