

TRADE AND COMMERCE  
CANADA

STANDARDS BRANCH

OTTAWA, May 30, 1960.

TYPE APPROVALMINNEAPOLIS-HONEYWELL BROWN  
"ELECTRONIK" RECORDERS WITH RETRANSMITTERS

The apparatus specified herein has been duly approved by the Standards Branch under the provisions of the Electricity Inspection Act, Chapter 94, R. S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Retransmitting slidewires for use with Models 153X10VH, 153X11VH and 153X12VH Strip Chart "Electronik" Recorders, manufactured by Honeywell Controls Limited, Toronto 17, Ontario.

Rating of Apparatus: See ratings for Models 153X10VH, 153X11VH and 153X12VH on Approval Circular S-EA.456, May 30, 1960.

Description: This circular grants approval for the use of retransmitting slidewires with Models 153X10VH, 153X11VH and 153X12VH Strip Chart Recorders. The slidewires used in the "Electronik" Recorders are of two types: the "straight" slidewire and the "helical" slidewire. When one type is used for measuring, the other type is used for retransmitting. Thus, in the cases of Models 153X10VH and 153X11VH, since the measuring slidewire is of the "helical" type, the retransmitting one is a 250-ohm "straight" type. The Model 153X12VH uses the "straight" type for measuring and a 250-ohm "helical" type is added when retransmitting is desired. The inclusion of a retransmitting slidewire is indicated in the model designation by the letters B5 so that a typical model number would be 153X10VH-X-20B5.

*E. F. Power*

E. F. Power,  
Assistant Director (E&G),  
Standards Branch.

*R. W. MacLean*  
R. W. MacLean,  
Director,  
Standards Branch.

