

SE-85-27
SL-100-359(J)
SL-100-SPE

Ottawa, November 25, 1966.

The Hydro-Electric Power Commission of Ontario,
620 University Avenue,
TORONTO 2, Ontario.

Attention: Mr. R. A. Thompson,
Protection and Control Engineer - Operations.

SPECIAL APPROVAL - Sixteen Only Honeywell "ElectroniK 15"
Two-Pen Potentiometer Recorders, Type
"SY153X(22)(VV) - x - (27)(A8)", Serial
Numbers 860537 to 860552 inclusive.

Dear Sirs:

This is to advise that Special Approval has been granted by the Standards Branch for use in Canada for billing purposes to sixteen Honeywell "ElectroniK 15" two-pen potentiometer recorders, type "SY153X(22)(VV) - x - (27)(A8)", serial numbers 860537 to 860552 inclusive.

These recorders have two 50-millivolt inputs, use chart number 963-0195, and have the standard cell, battery and standardizing rheostat replaced by a Zener diode regulated power supply.

These recorders will be used for billing purposes at various locations in Ontario by The Ontario Hydro-Electric Power Commission.

Yours very truly,

K. Cryer, Chief,
Electricity and Gas Division,
Standards Branch.

BAW/mee

c.c. to District Inspectors of Electricity and Gas at:
Belleville, Fort William, Hamilton, London,
Ottawa, Toronto and Sudbury, Ontario.