

SE-85-27
SL-100-359(L)
SL-100-SFE

Ottawa, March 21, 1967.

The Hydro-Electric Power Commission of Ontario,
620 University Avenue,
TORONTO 2,
Ontario.

Attention: Mr. R. A. Thompson,
Protection and Control Engineer - Operations.

SPECIAL APPROVAL - One Only Honeywell "ElectroniK 15"
Two-Pen Potentiometer Recorder,
Type 'SY153X(22)(VV)-x-(27)(A8)'
Serial Number 860545


Dear Sirs:

This is to advise that Special Approval has been granted by the Standards Branch for use in Canada for billing purposes to the above-named apparatus.

This recorder is one of a group of sixteen recorders that received special approval under Special Approval Letter of November 25, 1966 with 50 millivolts input for full scale on both channels and has now been changed so that one channel has a range 0-100 millivolts and the other a range 25-0-75 millivolts. (10)

This recorder is used by The Hydro-Electric Power Commission of Ontario in the Toronto District.

Yours very truly,


K. Gryer, Chief,
Electricity and Gas Division,
Standards Branch.

EAW/mce

c.c. to: Mr. W. R. McEown, District Inspector of Electricity and Gas, Toront