



Industry
Canada
Legal Metrology
Branch

Industrie
Canada
Direction de la
métrologie légale

Holland Avenue
Ottawa, Ontario
K1A 0C9

Avenue Holland
Ottawa (Ontario)
K1A 0C9

Telephone: (613) 952-0635

Téléphone: (613) 952-0635

Project: AP-GL-93-0054

Date: 1994-02-14

Mr. Ed Bons, P.Eng.
Rosemount Instruments Ltd.
808-55th Avenue, N.E.
Calgary, Alberta
T2E 6Y4

Dear Mr. Bons:

This is in response to your recent notification given in accordance with the Electricity and Gas Inspection Regulations, Section 13, regarding devices approved under Notices of Approval G-96-1, G-131, AG-0268 and AG-0269. It is noted that these devices have been tested for accuracy over the required ambient temperature range of -30°C to $+40^{\circ}\text{C}$, and found to be satisfactory. Consequently, it is not necessary to indicate the ambient temperature range on the nameplate.

Accordingly, your notification is acknowledged and the modifications are accepted as being included within the above-noted approval. All recipients of the approval notice are being informed by copy of this letter.

Yours truly,

H.L. Fraser, Chief
Electricity and Gas

cc: Approvals Mailing List