



Legal Metrology
Branch

Direction de la
Métrologie légale

Holland Avenue
Ottawa, Ontario
K1A 0C9

Avenue Holland
Ottawa, Ontario
K1A 0C9

Telephone: (613)952-0635

Your file *Votre référence*

Our file *Notre référence*

File No.: 06635
Project: AP-GL-93-0003
Date: 1993-04-21

Mr. Doug Warren
ITT Barton Instruments
3840 11A Street N.E.
Calgary, Alberta,
T2E 6M6

Dear Sir,


This is in response to your recent notification given in accordance with the Electricity and Gas Inspection Regulations, Section 13, regarding devices approved under Notices of Approval AG-0321 Rev. 1 and AG-0339. It is noted that these devices are to be modified as follows:

- To accommodate the use of added serial input ports for live input of gas analysis data from approved and compatible chromatographs, the following firmware versions have been developed by ITT Barton, and extend the list of approved firmware versions currently delineated in the above Notices of Approval.

- (i) N Gas G2.4.0V for the 1130 Scanner
- (ii) N Gas G2.4.0S for the 1110 and 1120 scanners
- N Gas G2.4.0C for the 1110 and 1120 scanners

I am advised by the Gas Metrology Laboratory that their evaluation of the proposed modifications indicates that they may be regarded as immaterial in relation to the approval provisions of the Act. Accordingly, your notification is acknowledged and the modifications are accepted as being included within the above-noted approval. All recipients of the approval notice are being informed by copy of this letter.

Yours truly,



H.L. Fraser, Chief
Electricity and Gas

cc: Approvals Mailing List



Consumer and
Corporate Affairs Canada

Consommation et
Affaires commerciales Canada

Legal Metrology
Branch

Direction de la
Métrologie légale

Holland Avenue
Ottawa, Ontario
K1A 0C9

Avenue Holland
Ottawa, Ontario
K1A 0C9

Téléphone: (613) 952-0635

Your file Votre référence

Our file Notre référence

Dossier: 06635
Projet: AP-GL-93-0003
Date: 1993-04-21

M. Doug Warren
ITT Barton Instruments
3840 - rue 11A, N.E.
Calgary, Alberta
T2E 6M6

Monsieur,

Nous accusons réception de l'avis de modification que vous nous avez envoyé récemment conformément à l'article 13 du Règlement sur l'inspection de l'électricité et du gaz, au sujet des appareils approuvés en vertu de l'avis d'approbation AG-0321 Rev. 1 et AG-0339. Il est à noter que ces appareils doivent être modifiés comme suit:

- Afin de permettre l'utilisation de portes en série additionnelles pour l'entrée directe des données d'analyse des gaz provenant des chromatographes approuvés et compatibles, les versions de microprogrammation suivantes ont été mises au point par ITT Barton et viennent s'ajouter à la liste des versions de microprogrammation approuvées et énumérées dans les avis d'approbation susmentionnés.
 - (i) N Gas G2.4.0V pour le lecteur 1130
 - (ii) N Gas G2.4.0S pour les lecteurs 1110 et 1120
 - N Gas G.2.4.0C pour les lecteurs 1110 et 1120

Après avoir étudié les modifications proposées, le Laboratoire de gaz (métrologie) me signale que ces modifications ne contreviennent en rien aux dispositions de la Loi. En conséquence, les modifications sont acceptées et intégrées de facto dans l'avis d'approbation susmentionné. Tous ceux qui ont reçu l'avis d'approbation sont avisés de cette décision par un copie de la présente lettre.

Je vous prie d'agréer mes salutations distinguées.


H.L. Fraser, Chef
Électricité et gaz

cc: Liste d'envoi des approbations

Canada

