



Industry Canada
Legal Metrology Branch

Industrie Canada
Direction de la métrologie légale

Holland Avenue
Ottawa, Ontario
K1A 0C9

Avenue Holland
Ottawa (Ontario)
K1A 0C9

Telephone: (613) 952-0639

Téléphone: (613) 952-0639

Project: AP-GL-94-0049

Date: 1995-01-18

ITT Barton Instruments
3840 - 11A Street, N.E.
Calgary, Alberta
T2E 6M6
Attn.: Andy Seeger

Dear Mr. Seeger,

This is in response to your recent notification given in accordance with the *Electricity and Gas Inspection Regulations*, Section 13 regarding devices approved under Notices of Approval AG-0321, AG-0321 Rev 1 and AG-0321 Rev 2. It is noted that these devices are to be modified as follows:

- Firmware version NGAS N2.3.5V is hereby approved for use in ITT Barton Scanner 1130 Electronic Volume Correctors under the above Notices of Approval.

I am advised by the Gas Metrology Laboratory that their evaluation of the proposed modifications indicates that they may be regarded as immaterial in relation to the approval provisions of the Act. Accordingly, your notification is acknowledged and the modifications are accepted as being included within the above-noted approval. All recipients of the approval notice are being informed by copy of this letter.

Yours truly,

H.L. Fraser, Director
Electricity and Gas

cc: Approvals Mailing List

Canada



Industry Canada
Legal Metrology Branch

Industrie Canada
Direction de la métrologie légale

Holland Avenue
Ottawa, Ontario
K1A 0C9

Avenue Holland
Ottawa (Ontario)
K1A 0C9

Telephone: (613) 952-0639

Téléphone: (613) 952-0639

Projet: AP-GL-94-0049
Date: 1995-01-18

ITT Barton Instruments
3840 - 11A Street, N.E.
Calgary, Alberta
T2E 6M6
Att.: Andy Seeger

Monsieur,

Nous accusons réception de l'avis de modification que vous nous avez envoyé récemment conformément à l'article 13 du Règlement sur l'inspection de l'électricité et du gaz, au sujet des appareils approuvés en vertu des avis d'approbation AG-0321, AG-0321 Rév. 1 et AG-0321 Rév. 2. Il est à noter que ces appareils doivent être modifiés comme suit :

- L'utilisation de la version NGAS N2.3.5V du microprogramme est approuvée dans les correcteurs de volume électroniques à lecteur 1130 de ITT Barton visés par les avis susmentionnés.

Après avoir étudié la modification proposée, le Laboratoire de gaz (métrologie) me signale qu'elle ne contrevient en rien aux dispositions de la Loi. En conséquence, la modification est acceptée et intégrée de facto dans les avis d'approbation susmentionnés. Tous ceux qui ont reçu les avis d'approbation sont avisés de cette décision par une copie de la présente lettre.

Je vous prie d'agréer mes salutations distinguées.

H.L. Fraser, Directeur
Électricité et gaz

c.c. : Liste d'envoi des approbations

Canada