



Consumer and
Corporate Affairs Canada

Consommation et
Affaires commerciales Canada

Legal Metrology
Branch

Direction de la
Métrologie légale

Holland Avenue
Ottawa, Ontario
K1A 0C9

Avenue Holland
Ottawa, Ontario
K1A 0C9

Your file Votre référence

Telephone: (613)952-0635

File No.: ^{Our file} 06635 ^{Notre référence}
Project: AP-GL-93-0017
Date: 1993-06-15

Mr. Doug Warren
ITT Barton Instruments
3840 - 11A Street, N.E.
Calgary, Alberta
T2E 6M6

Dear Mr. Doug Warren,

This is in response to your recent notification given in accordance with the Electricity and Gas Inspection Regulations, Section 13, regarding devices approved under Notice of Approval AG-0339, dated April 28, 1993. It is noted that these devices are to be modified as follows:

- i) The firmware versions listed in AG-0339 should have the "/" removed and a "0" field added; e.g. N/Gas B2.3C should be NGas B2.3.OC.
- ii) The following firmware versions are also approved:

NGas B2.4.OS	NGas B2.4.OC
NGas S2.4.OS	NGas S2.4.OC
NGas M2.4.OS	NGas M2.4.OC

I am advised by the Gas Metrology Laboratory that their evaluation of the proposed modifications indicates that they may be regarded as immaterial in relation to the approval provisions of the Act. Accordingly, your notification is acknowledged and the modifications are accepted as being included within the above-noted approval. All recipients of the approval notice are being informed by copy of this letter.

Yours truly,

H.L. Fraser, Chief
Electricity and Gas

cc: Approvals Mailing List

Canada