



Date: 2011-08-25

File number: AP-AG-11-0036

**MODIFICATION ACCEPTANCE LETTER (MAL)
FOR
Approved Meters and Devices**

Type of Device

Diaphragm, positive displacement gas meter

Purpose

The purpose of this letter is to convey details of modifications to gas measuring devices which have previously received approval of type recognition from Measurement Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered to be included within the scope of the originally issued Notices of Approval specified in this document and any subsequent revisions to these Notices of Approval.

Scope

The modifications described in this letter affect devices identified in the following Notice (s) of Approval:

Approval Number

G-0060, G-0094, G0138, G-0145, G-0101, AG-0292, AG-0310, AG-0315, AG-0373

Applicant

Sensus Metering Systems
805 Liberty Blvd, P.O. Box 528
DuBois, Pennsylvania, USA
15801

Description of Modifications

The Itron 100G ERT (AG-0546) automatic meter reader can be used with the Sensus meters in the attached list. The information contained in this MAL will be conveyed in the respective approvals for the meters identified when the approvals for these meters require revision.

Revision 1:

The purpose of revision 1 was to add to the following table the model RCM-750 natural gas diaphragm meter for use with the Itron model 100G DL Fixed Network (DLN) ERT automatic meter reader, part numbers ERG-5003-008 and ERG-5003-008P.



Date: 2011-08-25

File number: AP-AG-11-0036

Revision 2:

The purpose of revision 2 is to add to the following table the part numbers ERG-5000-008, ERG-5000-008P, ERG-5002-008 and ERG-5002-008P for the meter model RCM-750.

Itron 100G ERT, part number	Sensus meter models
ERG-5000-002, ERG-5002-002	RC-M-230, RC-230TC, R-275
ERG-5000-003, ERG-5002-003	MR-12TC, MR-8, MR-12
ERG-5000-004, ERG-5002-004	# 415TC, RC-415, # 415
ERG-5000-008, ERG-5002-008	RC-750, RC-1000, RC-1600, RC-3000, # 5000, # 10000
ERG-5000-008, ERG-5000-008P ERG-5002-008, ERG-5002-008P ERG-5003-008, ERG-5003-008P	RCM-750

Original copy signed by:
Randy Byrtus, CET
Manager, Gas Measurement



Date: 2011-08-25
Dossier: AP-AG-11-0036

LETTRE D'ACCEPTATION DE MODIFICATION (LAM)
pour
Compteurs et appareils approuvés

Type d'appareil

Compteur de gaz volumétrique, à membrane

Objet

La présente lettre vise à expliquer les détails des modifications apportées aux compteurs dont le type a déjà été approuvé par Mesures Canada. Ces modifications ont été évaluées par le Laboratoire des services d'approbation en conformité aux exigences nationales et doivent être considérées comme faisant partie de la portée des avis d'approbation d'origine énumérés dans le présent document et de toutes leurs révisions subséquentes.

Portée

Les modifications décrites dans la présente lettre s'appliquent aux compteurs visés par les avis d'approbation suivants :

Numéro d'approbation

G-0060, G-0094, G0138, G-0145, G-0101, AG-0292, AG-0310, AG-0315, AG-0373

Requérant

Sensus Metering Systems
805 Liberty Blvd, P.O. Box 528
DuBois, Pennsylvania, USA
15801

Description des modifications

Le modèle Itron 100G ERT de lecteur automatique de compteur (AG-0546) peut être utilisé avec les compteurs Sensus figurant sur la liste ci-jointe. Les informations contenues dans la présente LAM seront ajoutées à l'avis d'approbation respectif de chaque compteur mentionné, au moment de la révision de l'avis d'approbation.

Revisions 1:

Le but de la révision 1 était d'ajouter au tableau suivant le modèle RCM-750 compteur volumétrique de gaz naturel à membrane pour utiliser avec le modèle 100G DL Fixed Network (DLN) ERT lecteur automatique de compteur, numéros de la pièce ERG-5003-008 and ERG-5003-008P.



Date: 2011-08-25
Dossier: AP-AG-11-0036

Revisions 2:

Le but de la révision 2 est d'ajouter au tableau suivant les numéros de pièces ERG-5000-008, ERG-5000-008P, ERG-5002-008 et ERG-5002-008P pour le modèle de compteur RCM-750 .

Itron 100G ERT, numéro de la pièce	Modèles de compteurs Sensus
ERG-5000-002, ERG-5002-002	RC-M-230, RC-230TC, R-275
ERG-5000-003, ERG-5002-003	MR-12TC, MR-8, MR-12
ERG-5000-004, ERG-5002-004	# 415TC, RC-415, # 415
ERG-5000-008, ERG-5002-008	RC-750, RC-1000, RC-1600, RC-3000, # 5000, # 10000
ERG-5000-008, ERG-5000-008P ERG-5002-008, ERG-5002-008P ERG-5003-008, ERG-5003-008P	RCM-750

Copie authentique signée par:

Randy M. Byrtus, TEC
Gestionnaire, Mesure des Gaz