



Measurement Canada
An agency of Industry Canada

Mesures Canada
Un organisme d'Industrie Canada

MAL No. - NE De LAM
MAL-G122

Date: May 16, 2002

File number: AP-AG-02-0001

MODIFICATION ACCEPTANCE LETTER (MAL) FOR Approved Meters and Devices

Type of Device

Diaphragm Meter

Purpose

The purpose of this letter is to convey details of modifications to meters which have previously received approval of type recognition from Measurement Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered to be included within the scope of the originally issued Notices of Approval specified in this document and any subsequent revisions to these Notices of Approval.

Scope

The modifications described in this letter affect meters identified in the following Notice of Approval:

Approval Number

AG-0311

Applicant

Canadian Meter Company
275 Industrial Rd.
Cambridge Ontario
N3H 4R7

Description of Modifications

The Canadian Meter Company diaphragm meters: AL/ALM 800, AL/ALM 800C and the ALM/AL1000 now connect the front disk to the front diaphragm ring with an applied adhesive and a mechanical crimping process instead of soldering. Both methods of attachment are acceptable.

Original copy signed by:

Randy Byrtus
Approvals Technical Coordinator
Fluid Measurement Discipline



Measurement Canada
An agency of Industry Canada

Mesures Canada
Un organisme d'Industrie Canada

MAL No. - NE De LAM
MAL-G122

Date : 16 mai, 2002

Dossier : AP-AG-02-0001

LETTRE D'ACCEPTATION DE MODIFICATION (LAM)
pour
Compteurs et appareils approuvés

Type d'appareil

Compteur à membrane

Objet

La présente lettre vise à expliquer les détails des modifications apportées aux compteurs dont le type a déjà été approuvé par Mesures Canada. Ces modifications ont été évaluées par le Laboratoire des services d'approbation en conformité aux exigences nationales et doivent être considérées comme faisant partie de la portée de l'avis d'approbation d'origine énuméré dans le présent document et de toutes ses révisions subséquentes.

Portée

Les modifications décrites dans la présente lettre s'appliquent aux compteurs visés par l'avis d'approbation suivant :

Numéro d'approbation

AG-0311

Requérant

Canadian Meter Company
275 Industrial Rd.
Cambridge Ontario
N3H 4R7

Description des modifications

Sur les compteurs à membrane de la Canadian Meter Company des modèles AL/ALM 800, AL/ALM 800C et ALM/AL1000, le disque avant est désormais fixé à la bague de la membrane à l'aide d'un adhésif appliqué et d'un processus de sertissage et non plus par soudage. Les deux méthodes d'assemblage sont acceptables.

Copie authentique signée par:

Randy Byrtus
Coordonnateur des approbations techniques
Mesure des fluides