



Measurement Canada
An Agency of Industry Canada

Mesures Canada
Un organisme d'Industrie Canada

MAL No. - N° de LAM
MAL - G228

Date: May 23, 2007

File number: AP-AG-07-0001

MODIFICATION ACCEPTANCE LETTER (MAL) FOR Approved Meters and Devices

Type of Device

Electronic Volume Conversion Device

Purpose

The purpose of this letter is to convey details of modifications to meters which have previously received approval of type recognition from Measurement Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered to be included within the scope of the originally issued Notices of Approval specified in this document and any subsequent revisions to these Notices of Approval.

Scope

The modifications described in this letter affect meters identified in the following Notice (s) of Approval:

Approval Numbers

AG-0424, AG-0507, AG-0531C

Applicant

Romet Limited
1080 Matheson Blvd, East
Mississauga, Ontario
L4W 2V2

Description of Modifications

In addition to previously approved battery packs, the electronic volume conversion devices, the ECM2, the ECM2-PTZ and the ECM2-AT can be equipped with the alkaline battery pack, part number 34-068-0 with a cut off assembly. The cut-off assembly is intended to cut power to the electronic volume conversion device when the device experiences abnormal or sudden low voltage drops. This is intended to protect the electronic volume conversion device from "reverse" polarity connections.

Original signed by:

Randy Byrtus, CET
Technical Coordinator
Gas Measurement



Date : 23 mai 2007
Dossier : AP-AG-07-0001

LETTRE D'ACCEPTATION DE MODIFICATION (LAM) pour compteurs et appareils approuvés

Type d'appareil

Appareil de conversion du volume électronique

Objet

La présente lettre vise à expliquer les détails des modifications apportées aux compteurs dont le type a déjà été approuvé par Mesures Canada. Ces modifications ont été évaluées par le Laboratoire des services d'approbation en conformité aux exigences nationales et doivent être considérées comme faisant partie de la portée des avis d'approbation d'origine énumérés dans le présent document et de toutes leurs révisions subséquentes.

Portée

Les modifications décrites dans la présente lettre s'appliquent aux compteurs visés par l'avis d'approbation suivant :

Numéro d'approbation

AG-0424, AG-0507, AG-0531C

Requérant

Romet Limited
1080 Matheson Blvd, East
Mississauga, Ontario
L4W 2V2

Description des modifications

En plus des blocs-batteries approuvés précédemment, les appareils de conversion de volume électroniques, ECM2, ECM2-PTZ et ECM2-AT, peuvent être équipés d'un bloc-batterie alcalin, numéro de pièce 34-068-0 avec un mécanisme d'interruption. Le mécanisme d'interruption sert à couper l'alimentation de l'appareil de conversion de volume électronique en cas de baisse de tension soudaine ou anormale. Ce mécanisme a pour but de protéger l'appareil de conversion de volume électronique d'une polarité « inversée ».

Copie authentique signée par :

Randy Byrtus, T.S.A.I.
Coordonnateur technique
Mesure des gaz