



DEPARTMENT OF TRADE AND COMMERCE
STANDARDS BRANCH

G-13

OTTAWA October 20, 1965

NOTICE OF APPROVAL

FOR

CANADIAN METER COMPANY TYPE DU 5000 IRON CASE
POSITIVE DISPLACEMENT GAS METER

<u>Apparatus:</u>	Badged capacity	4,000 cu.ft./hr.(air)
	Differential pressure at badged capacity.	0.5" Water gauge
	Capacity per revolution.....	2.857 cu.ft.
	Working pressure.....	350 p.s.i.
	Tangent to 10 cu.ft. test dial revolution ratio.....	3.5 : 1
	Connections (female).....	4" N.P.T.

Description: The meter approved herein is identical to the type AL 5000 Gas Meter, approved in Circular S-GA 230, dated June 6th, 1962, except for the meter body. The cast aluminum body has been replaced with a ductile iron body that will withstand a higher range of working pressures.

For further details refer to the above mentioned circular.

Approval granted to: The Canadian Meter Company Limited,
Milton, Ontario and Edmonton, Alta.

for J.R. Gardiner
W. J. S. Fraser,
Chief, Standards Laboratory,
Standards Branch.

K. Cryer
K. Cryer,
Chief, Electricity & Gas Division,
Standards Branch.

Ref: SL:100:959B