



Consumer and
Corporate Affairs

Consommation et
corporations

Standards

Normes

NOTICE OF APPROVAL
AVIS D'APPROBATION

G-32-1

Ottawa, February 20, 1978

Rockwell, Type RC-225,
Standard And Temperature Compensated,
Positive Displacement Gas Meters

This Approval supplements Notice of Approval G-32, dated February 5, 1968, and covers the design modifications effected on this meter type by the manufacturer as described in the following.

Apparatus

The description of Rockwell Type RC-225 Standard and Temperature Compensated Positive Displacement Gas Meter is delineated under Notice of Approval G-32. With exception to the modifications described below, all details of the apparatus as outlined in the aforementioned Notice of Approval still apply.

Modification

The modified feature of the internal components is the molded, reinforced fiberglass, one piece, diaphragm port and pan. This replaces the diaphragm port and pan assembly made up of a plated steel diaphragm pan soldered to a white metal port.

This new design of molded port and pan is also being used in all Rockwell Type RC-230 gas meters approved under the Approval Notice G-117.

Approval granted to:

Rockwell International of Canada Limited,
Guelph, Ontario.

D. G. Smith, P.Eng.,
Chief, Electricity & Gas Division,
Legal Metrology Branch.

Ref: G 6635-R2-20



Consumer and
Corporate Affairs Canada

Consommation
et Corporations Canada

Legal Metrology
Branch

Direction de la
Métrologie légale

Holland Avenue
Ottawa, Ontario
K1A 0C9

Avenue Holland
Ottawa, Ontario
K1A 0C9

Your file *Votre référence*

Our file *Notre référence*

Telephone: (613)952-0639

File N°: 06635-N174
Project: AP-GL-89-0027

Northwestern Utilities Ltd.
10035-105th Street
EDMONTON, Alberta
T5J 2V6

1989-11-14

Attn.: Mr. Dennis Hutton

Dear Sir:

This is in response to your recent notification given in accordance with the Electricity and Gas Inspection Regulations, Section 13, regarding devices approved under Notices of Approval G-32, G117-2, and G125. It is noted that these devices are to be modified as follows:

- The RC225, RC230 and AL225 meters, approved under the above-stated notices of approval may now use handhole cover gaskets fabricated by Argus Industrial Supply, Winnipeg, Manitoba.

I am advised by the Gas Metrology Laboratory that their evaluation of the proposed modifications indicates that they may be regarded as immaterial in relation to the approval provisions of the Act. Accordingly, your notification is acknowledged and the modifications are accepted as being included within the above-noted approval. All recipients of the approval notice are being informed by copy of this letter.

Yours truly,

W.R. Virtue, Chief
Legal Metrology Laboratories

cc: Approvals Mailing List

Canada