



Department of consumer and corporate affairs / Ministère de la consommation et des corporations



STANDARDS BRANCH - DIRECTION DES NORMES

**NOTICE OF APPROVAL
AVIS D'APPROBATION**

G-38-1

OTTAWA April 11, 1973.

CANADIAN METER COMPANY BASE VOLUME INDEX, TYPE 3

This approval is supplementary to that of Circular G-38, dated January 30, 1969.

APPARATUS


Canadian Meter Company Base Volume Index, as covered under Approval Circular G-38. Except for the modifications described below, all details of the apparatus, as outlined in G-38, still apply.

DESCRIPTION OF MODIFICATIONS


1. The wriggler, connecting the BVI to the meter movement, is now cross-staked and not soldered.
2. The base plate shaft bearing material has been changed from bronze to Fluro-plastic.
3. Instead of nickel plated brass gears and pinions, black Celcon gears and aluminum pinions are introduced.
4. The top plate is positioned by two locating studs and is held down by two screws only.

Approval granted to:

Canadian Meter Company Ltd.,
Milton, Ontario
Edmonton, Alberta



J.S.T. Swanson, P. Eng.,
Chief, Standards Laboratory,
Standards Branch.



W.J.S. Fraser
Chief, Electricity & Gas Division
Standards Branch.

Ref: GL 1147-57/C6-187 (L)