



Department of consumer and corporate affairs / Ministère de la consommation et des corporations



STANDARDS BRANCH - DIRECTION DES NORMES

NOTICE OF APPROVAL

G-43-2

OTTAWA January 22, 1973.

MODIFICATION TO
SPRAGUE METER DIVISION, TYPES 175, 240 AND
250 STANDARD AND TEMPERATURE
COMPENSATED POSITIVE DISPLACEMENT
GAS METERS

Approval Circular G-43-1, dated February 24, 1972 remains in force.

DESCRIPTION

The modification involves the introduction of a plastic index in these meters in place of the metal one approved in earlier circulars.

The gear train assembly of the new index is made of a black plastic material, however, the front face is baked white ceramic on copper, whose appearance is similar to those on metal indexes.

The plastic index is directly interchangeable with the present metal version.

Approval granted to:

Sprague Meter Division of
Textron Canada Ltd.,
Hamilton, Ontario.

(for) J.S.T. Swanson, P. Eng.,
Chief, Standards Laboratory,
Standards Branch.

W.J.S. Fraser
Chief, Electricity & Gas Division,
Standards Branch.

Ref: GL-1147-57/S6-597P