



Department of consumer and corporate affairs / Ministère de la consommation et des corporations



STANDARDS BRANCH - DIRECTION DES NORMES

NOTICE OF APPROVAL

G - 86

OTTAWA January 24, 1972.

ROCKWELL TYPES "150", "210", "250", "310" and  
"415" POSITIVE DISPLACEMENT GAS METERS

Description

Approval has been granted by the Standards Branch for the elimination of a guard bar in the inlet port of Rockwell diaphragm meters, types "150", "210", "250", "310" and "415". The fitting of the guard bar was originally approved under Approval Notice SD-GA.148, dated November 25, 1958.

In other words these meters may now be supplied without any obstruction in the inlet port.

Approval granted to:

Rockwell Manufacturing Company,  
Guelph, Ontario.

J.S.T. Swanson, P. Eng.,  
Chief, Standards Laboratory,  
Standards Branch.

W.J.S. Fraser,  
Chief, Electricity and Gas Division,  
Standards Branch.

Re: SL-100-439