



Consumer and  
Corporate Affairs

Consommation et  
corporations

Standards

Normes

NOTICE OF APPROVAL  
AVIS D'APPROBATION

G-88-1

Ottawa, February 2, 1976.

MERCURY INSTRUMENTS INC. MODEL 99-C  
BATTERY OPERATED CHART DRIVE

This approval is supplementary to that of Circular G-88, dated May 30, 1972.

Description of Modification


This chart drive is identical to the chart drive approved in Circular G-88, except, that brass bushings replace the two jewels formerly used to support the escapement shaft.


The manufacturer recommends that the brass bushings be periodically cleaned and oiled to extend their life under adverse operating conditions.

All chart drives with the Serial No. D5000 and higher will include this modification.

This approval covers the use of this chart drive mechanism in any suitable, approved, circular chart recorder requiring a timing device.

Approval granted to: Parkinson Cowan (Canada) Ltd.,  
Chatham, Ontario.

  
J.L. Armstrong, P.Eng.

  
D.L. Smith, P.Eng.

Chief, Standards Laboratory, Chief, Electricity & Gas Division,  
Metrology and Laboratory Services.

Ref: GL 1147-57/M7-106

