



Consumer and
Corporate Affairs Canada

Consommation
et Corporations Canada

Legal Metrology
Branch

Direction de la
Métrologie légale

Holland Avenue
Ottawa, Ontario
K1A 0C9

Avenue Holland
Ottawa, Ontario
K1A 0C9

Telephone: (613) 990-8614

Your file *Voire référence*

Our file *Notre référence*

O7735-R292
October 1, 1985
AML-G-21

Romet Limited
1080 Matheson Blvd.,
Mississauga, Ontario
L4W 2V2
Attention: Mr. Harold J. Dunbar
Market Development Manager

Re: Application to Modify Approved Romet Limited Instrument Drive Rotary Gas Meters

Dear Mr. Dunbar:

This letter is in response to your letter dated September 12, 1985, regarding a modification to the type/models of meters described above, which are currently approved in Notice of Approval G-102-12 dated May 21, 1985. I am advised that the Gas Laboratory has evaluated the modification, of substituting a modified shear pin assembly made of molded plastic on instrument drive meters. The conclusion is that it may be regarded as insignificant, in relation to the approval provisions of the Gas Inspection Act.

Accordingly, pursuant to the authority delegated to me by the Director under section 7 of the Gas Inspection Regulations;

1. your letter is hereby acknowledged as satisfying the notification requirements of subsection 7(2) of the regulations; and
2. the proposed alteration is hereby accepted, in accordance with subsection 7(3) of the Regulations, as being an immaterial one and hence included within the approval.

All recipients of Notices of Approval under the Gas Inspection Act are being informed of this decision by copy of this letter. For our records you are requested to advise us of the starting serial number when the change is initiated in production.

Yours truly,

W.R. Virtue
Chief,
Legal Metrology Laboratories

cc: Mailing List (Gas Approvals)

Canada

