



**NOTICE OF APPROVAL
AVIS D'APPROBATION**

G-102-5

Ottawa, September 14, 1979

**ROMET LIMITED
SERIES RM 2000 ROTARY POSITIVE
DISPLACEMENT GAS METERS**

This approval supplements Notices of Approval G-102,
G-102-2 and G-102-3.

Apparatus

<u>Model</u>	<u>Max. Working Pressure psig</u>	<u>Max. Displacement cu. ft./hr</u>
RM 2000 Std. Digital Counter	175	2000
RM 2000 I.D.	175	2000
RM 2000 T.C.	175	2000
RM 2000 T.C./I.D.	175	2000

Description

The meters approved herein are identical in all respects to those previously approved except that nipple type connections are incorporated in the meter body as indicated in the photograph. The nipple is 1½ inch NPT by three (3) inches long. No change in identification or verification procedure is affected by this optional feature.

Approval granted to:

Romet Limited,
Mississauga, Ontario.

D.L. Smith,
Chief,
Electricity and Gas Division,
Legal Metrology Branch.

Ref: G-6635-R292

