



Consumer and
Corporate Affairs Canada

Consommation
et Corporations Canada

Legal Metrology
Branch

Direction de la
Métrologie légale

Holland Avenue
Ottawa, Ontario
K1A 0C9

Avenue Holland
Ottawa, Ontario
K1A 0C9

Your file *Votre référence*

Legal Metrology Laboratories
Telephone: (613) 952-0639

Our file *Notre référence*

O6635-D2
AP-GL-87-0009
July 14, 1987
AML-G-30

Dresser Canada Inc.
6688 Kitimat Road
Mississauga, Ontario
L5N 1P8
Attention: Mr. J. DeBenedictis

Dear Mr. DeBenedictis:

This is in response to your letter dated June 15, 1987, regarding Notice of Approval G-166 and proposing to modify the Model ETD Electronic Volume Corrector approved therein by:

- including a small backup battery for totalizer memory retention during primary battery replacement;
- consolidating on one basic board various circuits originally on separate boards; and
- changing the shape and size of the enclosure and making other minor changes, as documented.

I am advised that, upon evaluation of the proposed modifications, the Gas Laboratory has concluded that they may be regarded as insignificant in relation to the approval provisions of the Electricity and Gas Inspection Act. Accordingly, pursuant to the authority delegated to me under the said Act:

1. your letter is hereby acknowledged as satisfying the notification requirements of section 13, Part III of the Electricity and Gas Inspection Regulations, SOR/86-131, 16 January, 1986, and
2. the proposed alteration is hereby accepted, in accordance with the aforementioned section, as being an immaterial one and hence included within the approval.

All recipients of Notices of Approval under the Electricity and Gas Inspection Act are being informed of this decision by copy of this letter.

Yours truly,

W.R. Virtue
Chief

Canada

cc: Mailing List (Gas Approvals)