



Consumer and
Corporate Affairs Canada

Consommation
et Corporations Canada

Legal Metrology
Branch

Direction de la
Métrologie légale

Holland Avenue
Ottawa, Ontario
K1A 0C9

Avenue Holland
Ottawa, Ontario
K1A 0C9

Your file *Voire référence*

Our file *Notre référence*

Telephone: (613)952-0640

File N°: 06635
Project: AP-GL-92-0014

Mr. Schlebach
Daniel Industries Canada Inc.
4215-72nd Ave. S.E.
Calgary, Alberta
T2H 2H2

1992-06-08

Dear Sir,

This is in response to your recent notification given in accordance with the Electricity and Gas Inspection Regulations, Section 13, regarding devices approved under Notice of Approval G-196 dated 87-02-24. It is noted that these devices are to be modified as follows:

- In order to ensure centering of orifice plate in the fitting, the plate carrier has been changed by adding four (4) "protrusions" to it and, correspondingly, adding four (4) "notches" to the sealing ring (see attached sketch)

I am advised by the Gas Metrology Laboratory that their evaluation of the proposed modifications indicates that they may be regarded as immaterial in relation to the approval provisions of the Act. Accordingly, your notification is acknowledged and the modifications are accepted as being included within the above-noted approval. All recipients of the approval notice are being informed by copy of this letter.

Yours truly,

W.R. Virtue, Chief
Legal Metrology Laboratories

cc: Approvals Mailing List

Canada

PLATE CARRIER CHANGED BY ADDING 4 PROTRUSIONS AND
NOTCHING THE SEAL RING

