



Consumer and  
Corporate Affairs Canada

Consommation et  
Affaires commerciales Canada

Legal Metrology  
Branch

Direction de la  
Métrologie légale

Holland Avenue  
Ottawa, Ontario  
K1A 0C9

Avenue Holland  
Ottawa, Ontario  
K1A 0C9

*Your file*    *Votre référence*

Telephone: (613)952-0635

*Our file*    *Notre référence*

File N°: 06635  
Project: AP-GL-92-0053

Mr. Jan Knazko  
Romet Limited  
1080 Matheson Blvd., East  
Mississauga, Ontario  
L4W 2V2

1993-03-15


Dear Mr. Knazko,

This is in response to your recent notification given in accordance with the Electricity and Gas Inspection Regulations, Section 13, regarding devices approved under Notices of Approval G-102 through G-102-12, G-156, G-156-1, G-203 and AG-0316. It is noted that these devices are to be modified as follows:

- The thrust cover on these meters has been shortened slightly.

I am advised by the Gas Metrology Laboratory that their evaluation of the proposed modifications indicates that they may be regarded as immaterial in relation to the approval provisions of the Act. Accordingly, your notification is acknowledged and the modifications are accepted as being included within the above-noted approvals. All recipients of the approval notice are being informed by copy of this letter.

Yours truly,

  
 for H.L. Fraser, Chief  
 Electricity and Gas

*Vuong Nguyen*

cc: Approvals Mailing List



Consumer and  
Corporate Affairs Canada

Consommation et  
Affaires commerciales Canada

Legal Metrology  
Branch

Direction de la  
Métrologie légale

Holland Avenue  
Ottawa, Ontario  
K1A 0C9

Avenue Holland  
Ottawa, Ontario  
K1A 0C9

Your file    Votre référence

Téléphone: (613)952-0635

Our file    Notre référence

Dossier: 06635  
Projet: AP-GL-92-0053

1993-03-15

M. Jan Knazko  
Romet Limited  
6688 rue Kitimat  
Mississauga, Ontario  
L5N 1P8

Monsieur,

Nous accusons réception de l'avis de modification que vous nous avez envoyé récemment conformément à l'article 13 du Règlement sur l'inspection de l'électricité et du gaz, au sujet des appareils approuvés en vertu des avis d'approbation G-102 à G-102-12, G-156, G-156-1, G-203 et AG-0316, où il est indiqué que ces appareils doivent être modifiés comme suit:

- Le couvercle de butée de ces compteurs a été légèrement raccourci.

Après avoir étudié les modifications proposées, le Laboratoire de gaz (métrologie) me signale que ces modifications ne contreviennent en rien aux dispositions de la Loi. En conséquence, les modifications sont acceptées et intégrées de facto dans les avis d'approbation susmentionnés. Tous ceux qui ont reçu l'avis d'approbation sont avisés de cette décision par une copie de la présente lettre.

Je vous prie d'agréer mes salutations distinguées.

*Vuong Nguyen*

*for* H.L. Fraser, Chef  
Électricité et Gaz

cc: Liste d'envoi des approbations