



Consumer and
Corporate Affairs Canada

Consommation
et Corporations Canada

Legal Metrology
Branch

Direction de la
Métrologie légale

Holland Avenue
Ottawa, Ontario
K1A 0C9

Avenue Holland
Ottawa, Ontario
K1A 0C9

Your file Votre référence

Our file Notre référence

Telephone: (613)952-0640

File N°: 06635
Project: AP-GL-88-0060

Canadian Meter Co.
3037 Derry Road, West
Milton, Ontario
L9T 2X6
Attn.: Mr. G.F. Wilson

1989-04-03

Dear Mr. Wilson,

This is in response to your recent notification given in accordance with the Electricity and Gas Inspection Regulations, Section 13, regarding devices approved under Notices of Approval G-205, G-233 and G-238. It is noted that these devices are to be modified as follows:

- The incorporation of a modified software package, (identified as "software·pkg 1.2 (52699D001)"), expanding the display capability, on command, of information retained in the Electronic Volume Corrector.
- This modification does not, in any way, influence the calculation of corrected volume by the E.V.C.

I am advised by the Gas Metrology Laboratory that their evaluation of the proposed modifications indicates that they may be regarded as immaterial in relation to the approval provisions of the Act. Accordingly, your notification is acknowledged and the modifications are accepted as being included within the above-noted approval. All recipients of the approval notice are being informed by copy of this letter.

Yours truly,

W.R. Virtue, Chief
Legal Metrology Laboratories

cc: Approvals Mailing List

Canada